

MCV-5004 Aerospace Laser Calibration System with Dynamic Capability

System contents

L-109ER (Quantity of 2)

The LDDM laser head assembly is a compact device with a single aperture and extended range. A small retroreflector or a flat mirror can be used as target. The linear accuracy is traceable to NIST and meets ISO 17025 standard.

P-108E (Quantity of 1)

This processor module provide power supply, signal processing and automatic temperature and pressure compensation for two laser heads. Data is exported to a computer via an RS-232 or a PCMCIA interface.

ICTCP (Quantity of 1)

Air temperature and material temperature sensors; plug into the processor module. The barometric pressure sensor is located inside the processor module.

R-102 (Quantity of 2)

Two 0.5" (12.7 mm) diameter retroreflectors are provided as the targets on the moving axis or axes.

W-5000 (Quantity of 1)

Windows™ software for display, data collection and analysis of displacement errors, angular errors and straightness errors based on NMTBA, VDI, ISO, ASME B5.54, and ASME B5.57 standards is provided.

W-500SD (Quantity of 1)

Windows™ software for data collection and analysis of volumetric (vector) measurement is provided. This software also allows the operator to plot and view information on the axis displacement errors, vertical straightness errors, horizontal straightness errors and squareness errors. Compensation files for various controllers can also be generated as options.

W-500LB (Quantity of 1)

Windows™ software for high speed data collection and analysis of circular plots, velocity and acceleration is provided.

W-500QP (Quantity of 1)

Windows™ software for machine error diagnosis based on the laser/ballbar data is provided. This software analyzes the data and outputs information on various machine errors.

IPC5-1000 (Quantity of 2)

This PCMCIA card will provide high data rate up to 1000 data per second.

LD-03 (Quantity of 2)

Two magnetic bases are provided for mounting convenience.

LD-03P (Quantity of 2)

Two magnetic bases with post are provided for mounting convenience.

LD-14A (Quantity of 2)

Two adapter platforms are provided to hold the laser heads for spindle or machine mounting.

LD-37S (Quantity of 2)

Two 0.5" (12.7 mm) diameter beam benders are supplied to steer the laser beams for alignment convenience.

LD-58 (Quantity of 1)

A retroreflector mount for the body diagonal measurement.

LD-69/70 (Quantity of 2)

Two optical adapters (LD-69 is for short range and LD-70 is for long range) for flat-mirror targets.

LD-71 (Quantity of 2)

Two 6" (150 mm) flat mirrors are provided as the targets on the axes for circular contouring measurement.

LD-71S (Quantity of 1)

A 4"x3" (100 mm x 75 mm) flat mirror with steering is provided as the target on the axis for volumetric measurements.

LD-21L (Quantity of 2)

The system comes with 2 12-ft (3.6 m) cable sets for connecting the laser head and processor module. A 6-ft (1.8 m) power cord for connecting the processor module to line voltage, and a RS-232 cable are also provided.

LD-20D (Quantity of 2)

Two heavy-duty foam-lined carrying cases are supplied for easy transportation of the system.

SPECIFICATIONS:

Resolution	Linear: 1 microinch (0.01 micron) Angular: 1 microradian or 0.1 arcsec, typical Straightness: 1 microinch (0.01 micron), typical
Laser Frequency Stability	0.1 PPM
Measurement Accuracy	Linear: 1 PPM or ± 1 microinch/inch (± 1 micron/m) Angular: $\pm 0.2\%$ of reading
Range	100 feet (30 m) linear & angular 40 inches (1 m) circular 80"x60"x40" (2m x 1.5m x 1m) volumetric
Slew Rate	160 ips (4m/s), high slew rate available
Power	90-230 VAC, 50-60 Hz
Operating Environment	Temperature 60-90 °F (15-33 °C) Altitude 0-10,000 feet (0-3000m) Humidity 0-95% (non-condensing)



P-108E



W-5000



L-109ER
(Quantity of 2)



R-102
(Quantity of 2)



ICTCP



LD-14A
(Quantity of 2)



LD-70



LD-69



LD-71S



LD-37S
(Quantity of 2)



LD-58



LD-71
(Quantity of 2)



IPC5-1000
(Quantity of 2)



LD-03P
(Quantity of 2)

RS-232

AC CORD



LD-21L
(Quantity of 2)

LD-20D
(Quantity of 2)

